



## Proceedings

### PS01.01 | ePoster Session I

## Application Of Raman Microscopy For The Diagnosis Of The Chronic Lymphocytic Leukemia (CLL)

M. Fere<sup>\*1</sup>, L.H. Liu<sup>1</sup>, C. Gobinet<sup>1</sup>, A. Beljebbar<sup>1</sup>, V. Untereiner<sup>1</sup>, J.-F. Angiboust<sup>1</sup>, M. Manfait<sup>1</sup>, D. Gheldof<sup>2</sup>, H. Jacquemin<sup>2</sup>, S. Walbrecq<sup>2</sup>, E. Cornet<sup>3</sup>, X. Troussard<sup>3</sup>, B. Chatelain<sup>2</sup>, J. Angelo<sup>4</sup>, M. Chollat<sup>5</sup>, J. Klossa<sup>5</sup>, O. Piot<sup>1</sup>

<sup>1</sup>MEDyC CNRS UMR 7369, UFR de Pharmacie, Reims, France, <sup>2</sup>CHU Dinant Godinne, Namur, Belgium, <sup>3</sup>CHU Caen, Caen, France, <sup>4</sup>CMM-ARMINES, Fontainebleau, France, <sup>5</sup>TRIBVN, Châtillon, France

### *Introduction/ Background*

In hematology, actual diagnosis of B chronic lymphocyte-leukemia (CLL) is based on the microscopic analysis of cell morphology from patient blood smear. However, new photonic technologies appear promising to facilitate and improve the early diagnosis, prognostic and monitoring of personalized therapy. The development of automated diagnostic approaches could assist clinicians in improving the efficiency and quality of health services, but also reduce medical costs.

### *Aims*

The M3S project aims at improving the diagnosis and prognosis of the CLL pathology by developing a multimodal microscopy platform, including Raman spectrometry, dedicated to the automatic analysis of lymphocytes.

### *Methods*

Blood smears were prepared on glass slides commonly used in pathology laboratories for microscopy. Two types of sample per patient were prepared: a conventional blood smear and a deposit of "pure" lymphocyte subtypes (i.e. normal B, CLL B, T and NK), sorted out in flow cytometry by using the negative double labelling technique. The second sample is used for the construction of a database of spectral markers specific of these different cell types. The preparations were analyzed with the multimodal machine which combines i) a Raman micro-spectrometer, equipped with a 532nm diode laser excitation source; ii) a microscope equipped with 40x and 150x lenses and a high precision xyz motorized stage for scanning the blood smear, and localizing x-y coordinates of representative series (~100 for each patient) of lymphocyte cells before registering three Raman spectra; these cells of interest being previously localized by an original method based on the morphology analysis. After the Raman acquisitions, the conventional blood smears were submitted immunolabelling using specific antibodies. For the establishment of the Raman classifiers, this post-acquisition treatment was used as reference to distinguish the different lymphocyte sub-populations. Raman data were then analyzed using chemometric processing and supervised statistical classifiers in order to construct a spectral library of markers highly specific of the lymphocyte type and status (normal or pathological).

### *Results*

Currently, a total of 60 patients (CLL and healthy) were included in the study. Various classification methods such as LDA (Linear Discriminant Analysis), PLS-DA (Partial Least Square Discriminant Analysis), RF (Random Forest) and SVM (Support Vector Machine), were tested in the purpose to distinguish tumoral B lymphocytes from other cell types. These classification algorithms were combined with feature selection approaches. The best performances were around 70% of correct identification when a three-class model (B-CLL vs B-normal vs T and NK lymphocytes) was considered, and 80% in case of a two-class model (B-CLL vs B-normal



lymphocytes). These encouraging results demonstrate the potential of Raman micro-spectroscopy coupled to supervised classification algorithms for leukemic cell classification. The approach can find interest more generally in the field of cyto-hematology. Further developments will concern the integration of additional modality such as Quantitative Phase Imaging on one hand to speed the exploration process of cells of interest to be probed, and on the other hand to extract additional characteristics likely to be informative for CLL diagnosis. In addition, the identification of prognostic markers will be investigated by confronting the photonic data to clinical patient information.